CMA
230010349285

ilac-MRA CNAS 中国认可 国际互认 检测 TESTING CNAS L0619

# TEST REPORT

No.:（2026）wei song zi spectacles 5007

**Name of Product:** 1.56 Optical Resin Lens

**Client:** Danyang Shenhai Metal Materials Technology Co., Ltd.

**Entrusted by:** ————————————

**Manufactory:** Danyang Shenhai Metal Materials Technology Co., Ltd.

**Test Type :** Entrusted Test

## National Quality Inspection & Testing Centre for Ophthalmic Optics Products
## Danyang Inspection and Testing Center

# National Quality Inspection & Testing Centre for Ophthalmic Optics Products Test Report

No:(2026)wei song zi spectacles 5007 

| | | | |
|---|---|---|---|
| ***Name of Product** | 1.56 Optical Resin Lens | ***Specification** | See Remark |
| ***Sample Batch or Date of Production** | ———— | ***Trade Mark** | ———— |
| ***Client\Address\ Telephone\Postcode** | Danyang Shenhai Metal Materials Technology Co., Ltd.\No.18 Liyao Road,Danyang Development Zone,Danyang City,Zhenjiang City,Jiangsu Province,China\————\———— | | |
| ***Entrusted by\Address\Telephone\ Postcode** | ————\————\————\———— | | |
| ***Manufactory\Address\ Telephone\Postcode** | Danyang Shenhai Metal Materials Technology Co., Ltd.\No.18 Liyao Road,Danyang Development Zone,Danyang City,Zhenjiang City,Jiangsu Province,China\———\——— | | |
| **Test Type** | Entrusted Test | **Sum of Sampling** | YJWS20265007 |
| **Date received** | 2026-11-04 | **Sample number** | 8 pieces |
| **Test Date** | 2026-11-06 | **Sample Description** | Good，meet inspection requirements |
| **Inspection and Evaluation Criteria** | ISO 8980—2:2017 《Ophthalmic optics Uncut finished spectacle lenses—Part 2: Specifications for power-variation lenses | | |
| **Conclusions** | The tested items comply with the requirements of ISO 8980—2:2017. | | |
| | Date of Issued : 2026.11.11 | | |
| **Remark** | 1#～2#:0.00D ADD:+2.50D t:2.20mm φ70mm<br>Horizontal prism power:0.30△ Vertical prism power: 1.50△<br>3#～4#:0.00D ADD:+3.00D t:2.70mm φ70mm<br>Horizontal prism power:0.30△ Vertical prism power: 2.40△<br>5#～8#:-1.50D ADD:+2.75D t:2.20mm φ70mm<br>Horizontal prism power:0.50△ Vertical prism power: 1.80△<br>This report does not provide proof of compliance with the items not required to be tested by the client and/or functional indicators without the scope of reference marked on the products. | | |

Approver : Proofreader: Compiler:

# National Inspection & Testing Centre for Ophthalmic Optics Products Test Report

No.: (2026) wei song zi spectacles 5007



| No. | Standard clause | Item | | | Unit | Requirement | Test piece | Measurement | Comments |
|---|---|---|---|---|---|---|---|---|---|
| 1 | ISO 8980-2 5.2.2 | Optical require-ments | Tolerances on the focal power of lenses | Spherical power | $m^{-1}$ | ISO 8980-2 Table 1 | 8 | -0.10～+0.10 | Pass |
| | | | | Cylindrical power | | | | -0.08～+0.02 | Pass |
| 2 | ISO 8980-2 5.2.4 | | Variation power | | | ISO 8980-2 Table 3 | | 0.00～+0.10 | Pass |
| 3 | ISO 8980-2 5.2.5 | | prismatic power(△) | Horizontal component | cm/m | ISO 8980-2 Table 4 | | -0.12～+0.07 | Pass |
| | | | | Vertical component | | | | -0.17～+0.13 | Pass |
| 4 | ISO 8980-2 5.3 | Dimension | | Effective Dimension | mm | ≥-1 ≤+2 | | 0～+2 | Pass |
| | | | | Used Dimension | | ≥-2 | | 0～+2 | Pass |
| 5 | ISO 8980-2 5.3 | Tolerance on thickness | | | mm | ±0.3 | | -0. 2～+0. 2 | Pass |

Note : the information marked with "*" is provided by the customer ,and the customer is responsible for its authenticity.

————————Below Blank————————

# NOTICE

1. The report is invalid without a special seal for inspection and testing or the official seal of the inspection agency.

2. The copied report is invalid if it is not re-stamped with the special seal for inspection and testing or the official seal of the inspection agency.

3. The test report is invalid without the signatures of Main-inspector, Proofreader and Approver.

4. The test report is invalid if altered.

5. For "Entrusted Test," The testing data and results is applicable only to the submitted sample.

6. If the inspected entity has any objections to the test report, please raise your opinion to the organization of examination or the superior Market Supervision Administration within 15 days of receiving the report.

7. If the test result is qualified, the samples will be returned to the corporation after the period of opposition has expired. If not, the samples will be returned to the corporation after three months when the period of opposition has expired. If the corporation proposes not to take back the samples, then the corporation should confer with "The Testing Center" on how to deal with the samples.

---

# National Quality Inspection and Testing Centre for Ophthalmic Optics Products

National Quality Inspection and Testing Centre for Ophthalmic Optics Products (NOTC) was authorized by AQSIQ in June 2006. Its legal entity, Danyang Inspection and Testing Center, is a public institution directly under the Danyang Municipal government. All the test items passed the National Metrological Certification and Laboratory Assessment. NOTC has achieved global recognition for "one inspection, universally accepted" by signing international multilateral agreements through CNAS.

NOTC has a team of highly qualified staff, consisting of 26 professionals, including 3 professors, 4 senior engineers, 16 engineers, and 3 assistant engineers. The laboratory covers an area of 4700 m², with a constant temperature and humidity area of 60 m². NOTC owns 40 sets of domestic instruments and equipment, 35 sets of imported equipment, with a total value of RMB 24 million. All these equipment are at the advanced level. NOTC is one of the most authoritative national quality inspection and testing centers for ophthalmic optics products at home and abroad.

NOTC is the undertaking unit of the Secretariat of the Subcommittee 2 on Optometry Service of Technical Committee TC596 on Optometry of Standardization Administration of China (SAC/TC596/SC2), the Joint Secretariat of the sub Technical Committee for eye and face protection of the National Technical Committee for personal protection standardization (SAC/TC112/SC1), and the Secretariat of the Technical Committee for spectacle optical standardization of Jiangsu Province. It undertakes the preparation and revision of a number of national and local standards in the field of Optometry Service, and is responsible for the centralized management of national standards in professional and technical fields such as Optometry Service. It has outstanding advantages in standardization technology.

All testing items cover all the relevant National standards, Industry standards and ISO, EN, ANSI, BS, DIN and other relevant standards. NOTC has the ability to inspect on Spectacle frames, Optical glass blanks, Glass lenses, Resin lenses, Reading spectacle lenses, Mounted spectacle lenses, Photochromic lens blanks, Sunglasses, Contact lenses, Spectacles for driving, Personal Eye and Face Protection Devices.

To make the clients satisfied is our tenet. The quality with science and fairness is our core. NOTC will offer excellent service to the social and clients.

**Adress:YingChun Road 301 of DanYang city • Jiangsu province • China zip:212300**

**Business Tel(Fax):0511-86985158** **Supervision Tel:0511-86985058**

**Email:dyzjs5058@vip.163.com** **Http://www.cnotc.org**